

Certificate



Certificate no.

CU 72142585 01

License Holder:

Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
LITTLE ISLAND, CORK
Ireland

Manufacturing Plant:

Shenzhen WATT Electronics Co., Ltd.
No. 5, Tunnel 1, Tangfang Garden
35th District, Baoan County
518101 Shenzhen, Guangdong
China

Test report no.: USA-JAK 31482538 002

Client Reference: Diarmuid Hogan

Tested to:

UL 60950-1:2007 R12.11

CAN/CSA-C22.2 No.60950-1-07+A1:2011

Certified Product: Switched Mode Power Supply

License Fee - Units

Model Designation:

1) Xa1000-24defgh, 2) Xa500-24defgh

(a, d = 0-9, A-Z; e = "-", C, R, S;

f = 0-9, A-Z; g = 00-11; h = 0-9, A-Z, blank)

Rated Voltage: AC 100-240V, 50/60Hz

Rated Current: 1) 10-4.6A; 2) 4-2.3A

Protection Class: I

Output Ratings DC: 1) 19-28V/42A max.; 1008W max.

(derated at 5.94W/V below input AC 120V)

2) 19-28V/21A max.; 504W max.

(derated at 1.83W/V below input AC 180V)

Special Remarks:

Replaces Certificate CU72142065.

7

Appendix: 1, 1-12

7

Licensed Test mark:

Date of Issue

(day/mo/yr)

10/10/2014

